QUALITY CERTIFICATE

BATCH: 2491
PRODUCT: DILUCUP® ELEGANCE – BFD 9.0 mL
ARTICLE: 40DEBFD6
EXPIRY DATE: 30-JUL-2028

The expiry date is 30 months from production date when Dilucups are kept in the sealed aluminium package at +2°C to +30°C.
Once the aluminium package is opened Dilucups must be kept dry at +2°C to +8°C and be used within 3 months.

Medium formula
Monopotassium Phosphate 42.5 mg per liter deionized water*

*Adjusted and/or supplemented as required to meet performance criteria.

pH
Adjusted to 7.2 ± 0.2 at 25°C (according to FDA BAM).

Sterility
Sterility control shows < 1 CFU/mL after E-beam irradiation.

Volume control
Target volume is 9.0 mL.
Density at 20°C: 1.001 which means that 9.0 mL= 9.01 g.
Accepted deviation < ± 2% (according to FDA BAM).
Minimum measured deviation: -0.19 %
Maximum measured deviation: 0.46 %.

Productivity
Verified with E. coli strain WDCM 00013 and S. aureus strain WDCM 00034 (according to EN ISO 11133:2014).
Results for this batch: E. coli 14 %, S. aureus 8 %.

Volume, pH, sterility & productivity certified by:

February 16, 2026

Tove Svensson
Quality Manager

Warranty
Warranties arising in connection with the use of the products where such liability results from the actor omission of LabRobot Products AB shall be limited to the free replacement of the product(s) in question. For sterility guaranty, please refer to our ***Sterility Test Procedures in Dilucup Production.***